

Calibration Test Certificate

2018-08-27 09:23:28

Device

Serial Number:	KA416-1067969	Detector Type:	GasAlertMicroClip
Manufacturer:	BW Technologies	Next Cal Due:	2019-02-23

Test Result

Pass

Sensors

Type:	H2S	CO	LEL	O2
Result:	Pass	Pass	Pass	Pass
Final Reading:	25.0 ppm	100.0 ppm	50.0 %	20.9 %
Next Calibration Due:	2019-02-23	2019-02-23	2019-02-23	2019-02-23

Set Points

Type:	H2S	CO	LEL	O2
High Alarm:	15.0 ppm	200.0 ppm	10.0 %	23.5 %
Low Alarm:	10.0 ppm	25.0 ppm	5.0 %	19.5 %
TWA Alarm:	10.0 ppm	25.0 ppm		
STEL Alarm:	15.0 ppm	50.0 ppm		

Options

Datalog Interval:	15 seconds	Unit Programmed:	N/A
H2S STEL Period:	15 minutes	CO STEL Period:	15 minutes

Test Station

Dock Serial Number:	Z315-003295	Dock Location:			
	Inlet 1:	Inlet 2:	Inlet 3:	Inlet 4:	Inlet 5:
Used:	Yes	Yes	No	No	No
Concentration:	20.9 %	25.0	20.9 %	20.9 %	20.9 %
Type:	Purge	4 Gas Mixture 2.5% vol CH4	Purge	Purge	Purge
Notes:					